

Certificate of Analysis

HyServe
GmbH & Co. K.G.

Hechenrainer Str.24
82449 Uffing am Staffelsee

**Product
ID-No.
Lot Number
Expiry Date**

**Compact Dry VP for Vibrio spec.
1 000 900 / 1 001 014 / 1002 882
034011
30.04.2012**

☎ 0049-8846-1344
📠 0049-8846-1342
www.HyServe.com
info@hyserve.com

Test	Specifications	Results*
Appearance	Sheet of light yellow colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	8.6 –9.0	Corresponds
Colony growth		
<i>V. vulnificus</i> ATCC27562	Magenta colonies	Corresponds
<i>V. parahaemolyticus</i> ATCC17082	Blue/Blue-green colonies	Corresponds
<i>V. alginolyticus</i> 18	Milk white colonies	Corresponds
<i>V. cholerae</i> 11	Magenta colonies	Corresponds
<i>Pseudomonas aeruginosa</i> ATCC27853	No growth	Corresponds
<i>Enterococcus faecalis</i> ATCC19433	No growth	Corresponds
<i>Klebsiella oxytoca</i> ATCC13182	No growth	Corresponds

This lot meets the specifications.

Date: 10.01.11

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.